

**Certificate of conformity**

Dok. / Ref.	F 8.2.6.7
Leidimo Nr. / Issue No.	1.0
Data / Date	28/06/2024

Company: Esco Medical Technologies, UAB**Address:** Gamybos g. 2, Ramuciai, Kauno r., 54468 Lithuania**Name:** CultureCoin®**REF:** 1821493**LOT:** 030225

This certificate certifies that the product indicated above has been tested in accordance with the following test and meets its specifications.

Test results:

Description tests	Criteria	Result	Pass/Fail	Ref. No.
Visual inspection	WI 7.5.1.3	No defects	Pass	F 7.5.1.5
Limulus Amebocyte Lysate (LAL) test	<=20 EU/device	<0,500 EU/device	Pass	LAL.024.528.2025
Mouse Embryo Assay (MEA) test	average blastocyst formation rate >80%	95%	Pass	MEA-2025.02.01-00

Irradiation carried out according ISO 11137:2015; ISO 13485:2016 and MDR, 2017/745 requirements. Measured absorbed dose was from 28,8 kGy to 36,5 kGy.

Certificate issued (14/02/2025):

Quality engineer
Gediminas Zubras

